

آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
1 - 1	5.7	
2 - 1	5.6	
3 - 1	5.9	
4 - 1	5.5	
5 - 1	5.8	
6 - 1	5.9	
8 - 1	5.62	
9 - 1	5.6	
10 - 1	5.8	
11 - 1	5.7	
12 - 1	5.5	
13 - 1	5.8	
15 - 1	5.5	
17 - 1	5.7	
18 - 1	5.4	
19 - 1	5.7	
20 - 1	5.4	
22 - 1	5.8	
23 - 1	5.5	

Westgard	تفسیر	Hb A1C	آزمایش	Cobas اتوماتالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	5.6	SD	0.28	Exp. Date	1402/8/3	Lot. No	280801
Next Month	Mean	5.65	SD	0.16	CV	2.78			

