

آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	164	
12 - 1	163	
13 - 1	162	
15 - 1	163	
18 - 1	159	
19 - 1	158	
20 - 1	161	
22 - 1	162	
23 - 1	158	
24 - 1	155	
25 - 1	160	
26 - 1	163	
27 - 1	159	
29 - 1	158	
30 - 1	161	
31 - 1	163	

Westgard	تفسیر	ALP	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	160	SD	4.5	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	160.56	SD	2.53	CV	1.58			

