

# آزمایشگاه خیریه سلوژن همدان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	0.65	
12 - 1	0.63	
13 - 1	0.62	
15 - 1	0.59	
18 - 1	0.58	
19 - 1	0.59	
20 - 1	0.59	
21 - 1	0.59	
22 - 1	0.61	
24 - 1	0.58	
25 - 1	0.6	
26 - 1	0.58	
27 - 1	0.56	
29 - 1	0.57	
30 - 1	0.59	
31 - 1	0.58	

Westgard	تفسیر	Bili.D	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	0.59	SD	0.04	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	0.59	SD	0.02	CV	3.88			

