

آزمایشگاه خیره سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	8.5	
12 - 1	8.1	
13 - 1	8.5	
15 - 1	8.6	
18 - 1	8.5	
19 - 1	8.4	Cusum 3S
20 - 1	8	
22 - 1	8.5	Cusum 3S
23 - 1	8.1	Cusum 3S
24 - 1	8.5	Cusum 3S
25 - 1	8.4	Cusum 3S
26 - 1	8.5	Cusum 3S
27 - 1	8.2	Cusum 3S
29 - 1	8.1	Cusum 3S
30 - 1	8.3	Cusum 3S
31 - 1	8.4	Cusum 3S

Westgard	تفسیر	Calcium	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	8.3	SD	0.41	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	8.35	SD	0.19	CV	2.27			

