

آزمایشگاه خیره سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	125	
12 - 1	128	
13 - 1	122	
15 - 1	125	
18 - 1	127	
19 - 1	126	
20 - 1	127	
22 - 1	126	
23 - 1	125	
24 - 1	122	
25 - 1	127	
26 - 1	125	
27 - 1	122	
29 - 1	124	
30 - 1	130	
31 - 1	125	

Westgard	تفسیر	Cholesterol	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	126	SD	4.1	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	125.38	SD	2.22	CV	1.77			

