

آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	87	
12 - 1	84	
15 - 1	81	
18 - 1	88	
20 - 1	85	
22 - 1	90	
23 - 1	92	
25 - 1	87	
26 - 1	88	Cusum 3S
27 - 1	85	Cusum 3S
29 - 1	91	Cusum 3S
30 - 1	90	Cusum 3S
31 - 1	89	Cusum 3S

Westgard	تفسیر	Iron	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	86	SD	6.5	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	87.46	SD	3.1	CV	3.54			

