

آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	88	
12 - 1	90	
13 - 1	90	
15 - 1	88	
18 - 1	90	
19 - 1	91	
20 - 1	90	
22 - 1	89	
23 - 1	90	
24 - 1	90	
25 - 1	89	
26 - 1	90	
27 - 1	87	
29 - 1	91	
30 - 1	91	
31 - 1	91	

Westgard	تفسیر	Fasting Glucose	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	90	SD	2.55	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	89.69	SD	1.2	CV	1.33			

