

# آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	27	
12 - 1	28	
13 - 1	28	
15 - 1	27	
18 - 1	28	
19 - 1	28	
20 - 1	27	
22 - 1	28	
23 - 1	28	
24 - 1	28	1:3S
25 - 1	29	
26 - 1	27	
27 - 1	28	
29 - 1	28	
30 - 1	28	
31 - 1	29	

Westgard	تفسیر	HDL	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	28	SD	1.45	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	27.88	SD	0.62	CV	2.22			

