

آزمایشگاه خیریه سلوژن همدان

Control Serum - High Levey Jennings Chart

Day-Run	Value	Description
1 - 1	5.6	
2 - 1	5.7	
3 - 1	5.5	
4 - 1	5.6	
5 - 1	5.4	
6 - 1	5.6	
8 - 1	5.5	
9 - 1	5.6	
10 - 1	5.7	
11 - 1	5.5	
12 - 1	5.6	
13 - 1	5.4	
15 - 1	5.6	
18 - 1	5.7	
19 - 1	5.5	
20 - 1	5.4	
22 - 1	5.6	
23 - 1	5.5	

Westgard	تفسیر	Potassium (K)	آزمایش	الکترولیت آنالیزر	دستگاه	5	ماه	1401	سال
Control	Easy Lyte	Mean	5.5	SD	0.2	Exp. Date	1402/10/14	Lot. No	
Next Month	Mean	5.56	SD	0.1	CV	1.77			

