

# آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	289	
12 - 1	297	
13 - 1	291	
15 - 1	294	
18 - 1	290	
19 - 1	290	
20 - 1	291	
22 - 1	290	
23 - 1	293	
24 - 1	289	
25 - 1	291	
26 - 1	292	
27 - 1	289	
29 - 1	291	
30 - 1	290	
31 - 1	288	

Westgard	تفسیر	LDH	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	290	SD	4.5	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	290.94	SD	2.24	CV	0.77			

