

آزمایشگاه خیره سلوژن همدان

Control Serum - High Levey Jennings Chart

Day-Run	Value	Description
1 - 1	151	
2 - 1	152	
3 - 1	152	
4 - 1	151	
5 - 1	155	
6 - 1	154	
8 - 1	152	
9 - 1	155	
10 - 1	152	
11 - 1	155	
12 - 1	154	
13 - 1	155	
15 - 1	152	
18 - 1	156	
19 - 1	155	
20 - 1	151	
22 - 1	153	
23 - 1	154	

Westgard	تفسیر	Sodium (Na)	آزمایش	الکترولیت آنالیزر	دستگاه	5	ماه	1401	سال
Control	Easy Lyte	Mean	153	SD	3	Exp. Date	1402/10/14	Lot. No	
Next Month	Mean	153.28	SD	1.67	CV	1.09			

