

آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

| Day-Run | Value | Description |
|---------|-------|-------------|
| 11 - 1 | 35 | |
| 12 - 1 | 36 | |
| 13 - 1 | 35 | |
| 15 - 1 | 35 | |
| 18 - 1 | 36 | |
| 19 - 1 | 36 | |
| 20 - 1 | 35 | |
| 22 - 1 | 34 | |
| 23 - 1 | 35 | |
| 24 - 1 | 35 | |
| 25 - 1 | 34 | |
| 26 - 1 | 35 | |
| 27 - 1 | 35 | |
| 29 - 1 | 36 | |
| 30 - 1 | 34 | |
| 31 - 1 | 35 | |

| | | | | | | | | | |
|------------|-----------------|--------------|--------|----------------------|--------|-----------|-----------|---------|--------|
| Westgard | تفسیر | SGOT (AST) | آزمایش | mindray اتوآنالایزر- | دستگاه | 5 | ماه | 1401 | سال |
| Control | o MED DOUTROL M | Mean | 35 | SD | 1.5 | Exp. Date | 1401/5/13 | Lot. No | 220501 |
| Next Month | Mean | 35.06 | SD | 0.68 | CV | 1.94 | | | |

