

آزمایشگاه خیریه سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	3.5	
12 - 1	3.8	
13 - 1	3.9	
15 - 1	3.3	
18 - 1	4	
19 - 1	3.5	
20 - 1	3.6	Cusum 3S
22 - 1	3.6	Cusum 3S
23 - 1	3.5	Cusum 3S
24 - 1	3.5	
25 - 1	3.8	Cusum 3S
26 - 1	3.7	Cusum 3S
27 - 1	3.5	Cusum 3S
29 - 1	3.6	Cusum 3S
30 - 1	3.6	Cusum 3S
31 - 1	3.5	Cusum 3S

Westgard	تفسیر	Phosphorus (P)	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	3.6	SD	0.35	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	3.62	SD	0.18	CV	4.97			

