

آزمایشگاه خیره سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	50	
12 - 1	51	
13 - 1	49	
15 - 1	49	
18 - 1	49	
19 - 1	50	
20 - 1	49	
22 - 1	48	
23 - 1	51	
24 - 1	48	
25 - 1	47	
26 - 1	48	
27 - 1	48	
27 - 2	50	
28 - 1	48	
29 - 1	48	
30 - 1	49	
31 - 1	47	

Westgard	تفسیر	SGPT (ALT)	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	49.1	SD	2.5	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	48.83	SD	1.2	CV	2.46			

