

# آزمایشگاه خیره سلوژن همدان

## Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	89	
12 - 1	91	
13 - 1	90	
15 - 1	88	
18 - 1	85	Cusum 3S
19 - 1	90	Cusum 3S
20 - 1	89	Cusum 3S
22 - 1	86	Cusum 3S
23 - 1	91	Cusum 3S
24 - 1	85	Cusum 3S
25 - 1	89	Cusum 3S
26 - 1	85	Cusum 3S
27 - 1	90	Cusum 3S
29 - 1	92	Cusum 3S
30 - 1	89	Cusum 3S
31 - 1	91	Cusum 3S

Westgard	تفسیر	Triglycerides	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	87	SD	6.2	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	88.75	SD	2.32	CV	2.62			

