

# آزمایشگاه خیره سلوژن همدان

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
11 - 1	32	
12 - 1	33	
13 - 1	34	
15 - 1	33	
18 - 1	33	
19 - 1	34	
20 - 1	33	
22 - 1	33	
23 - 1	34	
24 - 1	33	
25 - 1	33	
26 - 1	34	
27 - 1	34	
29 - 1	34	
30 - 1	31	
31 - 1	31	

Westgard	تفسیر	Urea	آزمایش	mindray اتوآنالایزر-	دستگاه	5	ماه	1401	سال
Control	o MED DOUTROL M	Mean	33	SD	1.5	Exp. Date	1401/5/13	Lot. No	220501
Next Month	Mean	33.06	SD	1	CV	3.02			

